

Appendix for Hybrid Closed Loop Insulin Pump - Medtronic 670G/770G

Use this appendix in conjunction with the student's Diabetes Action and Management Plan.

STUDENT'S NAME

GRADE / YEAR

INSULIN DELIVERY

AUTO MODE



Blue shield on screen indicates the pump is in **AUTO** mode. The pump will alert the student that action is required when in AUTO mode.



Grey Shield - Safe Basal becomes active when the pump requires more information e.g. a Blood Glucose Level (BGL) check.

Auto mode / Safe basal means basal insulin delivery is informed by a continuous glucose monitor (CGM). The student **MUST** respond to all pump commands. If the pump request has not been attended to, it will switch to **MANUAL** mode.



MANUAL MODE


No Blue or Grey shield seen on screen means the pump has switched to **MANUAL** mode. Insulin is still being safely delivered.

EATING AND DRINKING

In auto mode: The student enters their finger prick **BLOOD** glucose reading and **MUST** bolus for all carbohydrate food, except for hypo treatment.


In manual mode: The student enters their finger prick **BLOOD** glucose reading and **MUST** bolus for all carbohydrate food, except for hypo treatment and activity food (if required).

PHYSICAL ACTIVITY



DISCONNECTING PUMP BEFORE SPORT

Select '**Suspend Delivery**' from the main menu by pressing the 'select' button and scrolling down to the **SUSPEND DELIVERY** option.



RECONNECTING PUMP AFTER SPORT

Reconnecting pump select '**Resume delivery**' from the main menu by pressing the 'select' button and scrolling down to the **RESUME DELIVERY** option.