

## Setting up Humidity in the Hybrid Incubator (BabyLeo)

Attach a water bag with the connection kit via the Leur lock on the rear sensor wall (see picture).



Touch the humidity therapy control knob. The therapy control turns yellow. Set the desired humidity value manually using the rotary knob.

The current humidity level in the patient compartment shows on the monitoring area on the main screen.

